

Pall (Canada) Ltd.  
3450 Ridgeway Drive  
Mississauga, ON L5L 0A2 Canada  
905-542-0330 phone  
www.pall.com

August 1, 2018

Shawn Glover  
Industrial Plankton Inc  
3378 Burns Avenue  
Victoria, BC V8Z 3P2

Dear Mr. Shawn Glover,

This letter is in response to your request for integrity test values for Pall Kleenpak™ capsules containing Ultipor® N66 membrane filter(s) (P/N KA3NTP6), using water as a wetting fluid. These values apply to a test temperature of  $20 \pm 5^\circ\text{C}$ . During the test, the temperature should not vary by more than  $1^\circ\text{C}$ .

**Forward Flow Test Parameters using Air**

Filter Part Number	Wetting Fluid	Test Pressure	Maximum Allowable Forward Flow
KA3NTP6	WATER	75 psig	5.3 cc/min

Pall Corporation is pleased to have assisted Industrial Plankton Inc with this request. If I can be of further assistance, do not hesitate to contact me at 905-593-4362.

Sincerely Yours,

Angelica Zdunski

Technical Specialist -  
Biotech